


Product Selection Decision Tree

Sample Risk Assessment Indicators:

- Limited ability to reposition self
- Braden score of 18 or less
- Braden Activity subscore of 1-3
- Braden Mobility subscore of 1 or 2
- Braden Friction and Shear subscore of 1 or 2
- Existing or history of pressure injuries



- Perform a hand check every shift and adjust inflation as needed
- Turn and reposition the patient every 2 hours or per individualized turning schedule
- Use only 1 incontinence pad at a time
- Send WAFFLE Overlay with patient at discharge